

Now with new parameters
RDW-SD
P-LCR

MYTHIC™ 22 OT

24 parameters haematology analyzer Open Tube system



Orphée

Mythic™ 22 OT

Technical specifications

Throughput	50 samples per hour
Sample volume	< 17 µl
Technology	• Impedance (WBC, RBC, PLT) • Spectrophotometry (HGB) • Optical (5-DIFF): Optical laser method for 5-part WBC differential using semiconductive diode
24 parameters	WBC, LYM% and #, MON% and #, NEU% and #, BAS% and #, EOS% and #, RBC, HGB, HCT, MCV, MCH, MCHC, RDW, PLT, MPV, PCT, PDW, RDW-SD, P-LCR
Curves	RBC curve, PLT curve
Scattergram	5-Diff scattergram
Dimensions	250 x 340 x 350 mm (W x D x H)
Weight	11 kg
Power	• 90 to 250 V • 50 to 60 Hz • Less than 150 VA
External printers	• Dot matrix printer • Monochrome inkjet printer, PCL3 type • Colour laser printer, PCL6 type
User's interface	• TFT colour touchscreen • Built-in numerical keyboard
Output	• RS232 • Ethernet (UDP/IP, TCP/IP protocol) • 2 USB ports • Centronics port • Barcode scanner port

Software specifications

Multilanguage	Brazilian Portuguese, Bulgarian, Chinese, Deutsch, English, French, Magyar, Polish, Portuguese, Romanian, Russian, Spanish
Calibration	Automatic or manual
Quality control	• 6 batches: 100 runs/batch • Table and Levey Jennings graphs
Archives	• Internal memory: 1500 patients • Mass storage (USB): up to 60 000 patients
Traceability	• Patient identification: name, patient ID, sample ID • Daily log • Reagents monitoring: lot number, expiry date, volume • Service log
Flagging	• Pathologic flags • Lab limits • Reagent alerts • Instrument alerts
Options	External barcode scanner

Reagents	
HM 22-001-1	• Cleaner for Mythic™ 22
HM 22-002-1	• OnlyOne™
HM 22-003-10	• Diluent for Mythic™ 22
Calibrator	Myt-Cal: calibrator for Mythic™ 22
Controls	Myt-5D tri level or normal level: quality controls for Mythic™ 22



24 parameters haematology analyzer Open Tube system

Orphée reserves own rights to technical and software modification.
In case of printer choice please contact your local Orphée's representative.

Specifications are correct at the time of printing: 09/2016 ver. 01